

PIN

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Cher Snyder, Assistant DEO Engineering & Compliance	LABORATORY NO	<u>1602109</u>
		DATE RECEIVED	<u>01/21/2016</u>
SAMPLE DESCRIBED AS	One Summa Canister, 24 h sample	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
SAMPLING LOCATION	Porter Ranch Community Elementary School (PRCS) Porter Ranch, CA 91326	ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54058</u>
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	6
CO ₂ , ppm	431
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 3/3/16

Approved By: 

Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1602109</u>
One Summa Canister, 24 h sample	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	10.33	-0.64	PASS
CH ₄ , ppmvC	0.3	10.17	10.65	4.77	PASS
CO ₂ , ppmvC	0.4	10.38	10.50	1.13	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	10.34	-5.99	PASS
NM/NEOC, ppmvC	0.2	10.64	10.24	-3.73	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	10069.42	-0.30	PASS
CH ₄ , ppmvC	0.3	9950	10387.01	4.39	PASS
CO ₂ , ppmvC	0.4	10100	10012.72	-0.86	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	10212.54	2.74	PASS
NM/NEOC, ppmvC	0.2	10000	10132.38	1.32	PASS

DATE ANALYZED 02/02/2016
REFERENCE NO: 16QM2AA
 QM2-101-70

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVOICE
 LAP AUDIT
 LABORATORY

WO #: 1602109



TO: SCAQMD LAB: OTHER:
 SOURCE NAME: Southern California Gas Co. I.D. No. _____
 Source Address: 12801 Tampa Ave City: Porter Ranch
 Mailing Address: _____ City: _____ Zip: 91326
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 1/21/16
 Approved by: Jason Low Office: _____ Budget #: 44716
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other

Sample Collected by: Robert Wimmer Date: 1/20/16 Time: 12:00

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch Community Elementary School (PRCS)	54058	1/19/16 / 00:00 / 24 hours	<-30"	+11.5

Relinquished by	Received by	Firm/Agency	Date	Time
R Wimmer <i>[Signature]</i>	<i>[Signature]</i>	SCAQMD Lab	1/21/16	7:40

Remarks: 1:3 scheduled samples from trailer at PRCS
 Porter Ranch Community School Elementary – 5 Sesnon Blvd, Porter Ranch, CA 91326
 GPS (34.293369, -118.580505)