

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2


<b>TO</b>	Jason Low Atmospheric Measurements Manager Science and Technology	<b>LABORATORY NO</b>	<u>1610327</u>
<b>SAMPLE DESCRIBED AS</b>	24 hour sample Canister # 53387	<b>DATE RECEIVED</b>	<u>04/13/2016</u>
<b>SAMPLING LOCATION</b>	Porter Ranch Castlebay Elementary School	<b>FACILITY ID NO</b>	<u>NA</u>
		<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
		<b>ST NO / PROJECT</b>	<u>NA</u>

---

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

Type	Canister
Number	<u>53387</u>
Pressure (Torr)	1400
CO, ppm	< 3
CH <sub>4</sub> , ppm	2
CO <sub>2</sub> , ppm	429
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 4/15/16

Approved By: 

Solomon Teffera, Acting Senior Manager  
Laboratory Services  
909-396-2199

*Sumner Wilson*  
*4-14-16*

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1610327</u>
24 hour sample Canister # 53387	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**


**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.14	11.38	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	2.27	12.37	PASS
CO <sub>2</sub> , ppmvC	0.3	1.57	2.03	29.11	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	1.99	-1.91	PASS
NM/NEOC, ppmvC	0.4	2.03	2.30	13.35	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1062	2.54	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1056	-1.09	PASS
CO <sub>2</sub> , ppmvC	0.3	1022	1044	2.19	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1046	0.23	PASS
NM/NEOC, ppmvC	0.4	1024	1021	-0.21	PASS

<b>DATE ANALYZED</b>	04/14/2016
<b>REFERENCE NO:</b>	16QM2AB
	QM2-101-98

WO #: 1610327



**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

DISTRICT  
 INVOICE  
 LAP A  
 LABORATORY NO.

TO: SCAQMD LAB:  OTHER:  \_\_\_\_\_  
 SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_  
 Source Address: 12801 Tampa Ave City: Porter Ranch  
 Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326  
 Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 4/13/16  
 Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716  
 REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill   
 Suspected Violation Rule(s) \_\_\_\_\_ Other  \_\_\_\_\_

Sample Collected by: Qian Zhou Date: 4/13/16 Time: 10:30am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	53387	4/12/16 / 00:00 / 24 hours	-30"	+22

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>[Signature]</i>	SCAQMD Lab	4/13/16	10:15

Remarks: 1:3 scheduled samples from trailer at Castlebay  
 Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA