

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low, Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1610920</u>
		DATE RECEIVED	<u>04/19/2016</u>
SAMPLE DESCRIBED AS	24 hour sample Canister # E3843	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
SAMPLING LOCATION	Porter Ranch Castlebay Elementary School	ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>E3843</u>
Pressure (Torr)	1351
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	423
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 4/27/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

4/27/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1610920</u>
24 hour sample Canister # E3843	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.13	10.49	PASS
CH ₄ , ppmvC	0.3	2.02	2.16	6.65	PASS
CO ₂ , ppmvC	0.3	1.57	2.13	35.89	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.18	7.63	PASS
NM/NEOC, ppmvC	0.4	2.03	2.22	9.31	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1063	2.60	PASS
CH ₄ , ppmvC	0.3	1068	1055	-1.19	PASS
CO ₂ , ppmvC	0.3	1022	1047	2.46	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1049	0.44	PASS
NM/NEOC, ppmvC	0.4	1024	1027	0.29	PASS

DATE ANALYZED	04/19/2016
REFERENCE NO:	16QM2AB
	QM2-101-99

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DI
 IN
 L
LAB



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 4/19/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 4/19/16 Time: 09:25am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	E3843	4/18/16 / 00:00 / 24 hours	-30"	+11.5

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhou Qian</i>	<i>JL</i>	SCAQMD Lab	<i>4/19/16</i>	<i>12:51</i>

Remarks: 1:3 scheduled samples from trailer at Castlebay
Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA