

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2


TO	Jason Low ADEO Science and Technology Advancement	LABORATORY NO	<u>1626736</u>
SAMPLE DESCRIBED AS	Triggered sample Canister 54041	DATE RECEIVED	<u>09/23/2016</u>
SAMPLING LOCATION	Highlands Community Center	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54041</u>
Pressure (Torr)	760
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	408
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 9/28/16

Approved By: _____


Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1626736</u>
24 hour sample Canister 54041	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.26	17.21	PASS
CH ₄ , ppmvC	0.3	2.02	1.94	-3.91	PASS
CO ₂ , ppmvC	0.4	1.57	6.27	299.59	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.11	4.11	PASS
NM/NEOC, ppmvC	0.5	2.03	2.20	8.71	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1038	0.19	PASS
CH ₄ , ppmvC	0.3	1068	1035	-3.08	PASS
CO ₂ , ppmvC	0.4	1022	1021	-0.11	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1022	-2.12	PASS
NM/NEOC, ppmvC	0.5	1024	995	-2.81	PASS

DATE ANALYZED	09/23/2016
REFERENCE NO:	16QM2AD
	QM2-101-122

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT INFORMATION

1626736



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 9/23/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 9/23/16 Time: 13:05

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54041	9/23/16 / 06:50 / 5 min	-30"	-2"

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhou</i>	<i>Q. Zhou</i>	SCAQMD Lab	09/23/16	1750

Remarks: Samples collected by passive sampling via the XonTech 912 triggered by the Thermo 55i NMHC.

Trigger is set to 5ppm

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)